

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL031921\
 Data File : VL036459.D
 Acq On : 19 Mar 2021 23:25
 Operator :
 Sample : OC031821 CAN 10045
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC031821 CAN 10045

Manual Integrations
 APPROVED

MMDadoda
 3/24/2021 6:14:55 PM

Quant Time: Mar 20 01:06:54 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL030321AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Mar 0
 QLast Update : Wed Mar 03 14:18:24 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.64	49	1594097	10.00	ppbv	-0.04
33) 1,4-Difluorobenzene	7.14	114	3591102	10.00	ppbv	-0.04
55) Chlorobenzene-d5	12.05	117	3293548	10.00	ppbv	-0.04
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.38	95	2658430	9.82	ppbv	-0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.20%
Target Compounds						
9) Propene	3.00	41	10172m	0.124	ppbv	Ovalue
15) Acetone	3.85	43	66718m	0.366	ppbv	
20) Methylene Chloride	4.32	84	31674	0.516	ppbv #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL031921\
Data File : VL036459.D
Acq On : 19 Mar 2021 23:25
Operator :
Sample : OC031821 CAN 10045
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC031821 CAN 10045

Manual Integrations
APPROVED

MMDadoda
3/24/2021 6:14:55 PM

Quant Time: Mar 20 01:06:54 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL030321AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Mar
QLast Update : Wed Mar 03 14:18:24 2021
Response via : Initial Calibration

