


Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL031921\
 Data File : VL036447.D
 Acq On : 19 Mar 2021 13:48
 Operator :
 Sample : BFB
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL030321AIR.M
 Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Wed Mar 03 14:18:24 2021
 Last Update : Wed Mar 03 14:18:24 2021



AutoFind: Scans 3609, 3610, 3611; Background Corrected with Scan 3585

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	25.9	271957	PASS
75	95	30	66	60.0	630528	PASS
95	95	100	100	100.0	1050517	PASS
96	95	5	9	7.2	75309	PASS
173	174	0.00	2	0.8	5146	PASS
174	95	50	120	62.4	655509	PASS
175	174	4	9	8.0	52541	PASS
176	174	93	101	95.0	622954	PASS
177	176	5	9	7.0	43296	PASS