

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL032119\
Data File : VL033424.D
Acq On : 21 Mar 2019 20:36
Operator : SY/AP
Sample : OC032019 CAN 10297
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC032019 CAN 10297

Quant Time: Mar 22 15:34:29 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL031519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Mar 15 14:52:14 2019
QLast Update : Fri Mar 15 14:52:14 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	879274	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	2021701	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.24	117	1885851	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.57	95	1481827	10.60	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	106.00%

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL032119\
Data File : VL033424.D
Acq On : 21 Mar 2019 20:36
Operator : SY/AP
Sample : OC032019 CAN 10297
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC032019 CAN 10297

Quant Time: Mar 22 15:34:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL031519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Mar 15 14:52:14 2019
QLast Update : Fri Mar 15 14:52:14 2019
Response via : Initial Calibration

