

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032122\
 Data File : VL038656.D
 Acq On : 21 Mar 2022 18:56
 Operator : SY/AP
 Sample : VL0321ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0321ABL01

Quant Time: Mar 22 03:50:02 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL032122AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Mar 21 17:09:26 2022
 QLast Update : Mon Mar 21 17:09:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.665	49	937696	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	7.169	114	2359581	10.000	ppbv	0.00
55) Chlorobenzene-d5	12.079	117	2020121	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.407	95	1568118	10.125	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032122\
Data File : VL038656.D
Acq On : 21 Mar 2022 18:56
Operator : SY/AP
Sample : VL0321ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0321ABL01

Quant Time: Mar 22 03:50:02 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL032122AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Mar 21 17:09:26 2022
QLast Update : Mon Mar 21 17:09:26 2022
Response via : Initial Calibration

