

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032122\
 Data File : VL038670.D
 Acq On : 22 Mar 2022 3:56
 Operator : SY/AP
 Sample : N1964-13DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 C0P83DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin
 Yesilyurt

03/22/2022

Supervised By :Mahesh
 Dadoda

03/24/2022

Quant Time: Mar 22 05:09:56 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL032122AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Mon Mar 2
 QLast Update : Mon Mar 21 17:09:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.658	49	935230	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	7.163	114	2425595	10.000	ppbv	0.00
55) Chlorobenzene-d5	12.073	117	2136792	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.404	95	1542485	9.415	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	94.200%
Target Compounds						
					Qvalue	
9) Propene	3.022	41	9465	0.198	ppbv	95
13) Ethanol	3.613	45	10593m	2.014	ppbv	
15) Acetone	3.851	43	41311	0.512	ppbv #	86
20) Methylene Chloride	4.337	84	80079	2.356	ppbv	94
26) Hexane	5.681	57	34765	0.382	ppbv #	40
34) 2-Butanone	5.247	43	36820m	0.384	ppbv	
53) Toluene	9.793	91	63349	0.315	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032122\
Data File : VL038670.D
Acq On : 22 Mar 2022 3:56
Operator : SY/AP
Sample : N1964-13DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
C0P83DL

Manual Integrations
APPROVED

Reviewed By : Semsettin
Yesilyurt

03/22/2022
Supervised By : Mahesh
Dadoda

03/24/2022

Quant Time: Mar 22 05:09:56 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL032122AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Mar
QLast Update : Mon Mar 21 17:09:26 2022
Response via : Initial Calibration

