

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032122\
 Data File : VL038672.D
 Acq On : 22 Mar 2022 5:15
 Operator : SY/AP
 Sample : N1964-14DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 C0P84DL

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 03/22/2022
 Supervised By : Mahesh Dadoda 03/24/2022

Quant Time: Mar 22 06:47:14 2022

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL032122AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Mar 22 2022

QLast Update : Mon Mar 21 17:09:26 2022

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.655	49	932443	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	7.160	114	2351046	10.000	ppbv	0.00
55) Chlorobenzene-d5	12.069	117	2039468	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.401	95	1476216	9.441	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	94.400%
Target Compounds						
					Qvalue	
9) Propene	3.018	41	7626	0.160	ppbv	91
13) Ethanol	3.604	45	7936m	1.513	ppbv	
15) Acetone	3.854	43	19145m	0.238	ppbv	
20) Methylene Chloride	4.337	84	31635	0.933	ppbv	96
26) Hexane	5.681	57	14220m	0.157	ppbv	
31) cis-1,2-Dichloroethene	5.539	61	9441m	0.111	ppbv	
53) Toluene	9.793	91	30528	0.157	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032122\
Data File : VL038672.D
Acq On : 22 Mar 2022 5:15
Operator : SY/AP
Sample : N1964-14DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
C0P84DL

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 03/22/2022
Supervised By : Mahesh Dadoda 03/24/2022

Quant Time: Mar 22 06:47:14 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL032122AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Mar
QLast Update : Mon Mar 21 17:09:26 2022
Response via : Initial Calibration

