

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032122\
 Data File : VL038675.D
 Acq On : 22 Mar 2022 7:24
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 03/23/2022
 Supervised By :Mahesh Dadoda 03/24/2022

Quant Time: Mar 23 03:41:04 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL032122AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Mar 2
 QLast Update : Mon Mar 21 17:09:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.652	49	862101	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	7.156	114	2170309	10.000	ppbv	0.00
55) Chlorobenzene-d5	12.063	117	1862897	10.000	ppbv	-0.01
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.394	95	1423818	9.969	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.700%
Target Compounds						
					Qvalue	
15) Acetone	3.870	43	8856m	0.119	ppbv	
32) 1,1,1-Trichloroethane	6.491	97	5158m	0.040	ppbv	
35) Carbon Tetrachloride	7.009	117	4757m	0.038	ppbv	
52) Tetrachloroethene	11.208	164	1860m	0.030	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032122\
Data File : VL038675.D
Acq On : 22 Mar 2022 7:24
Operator : SY/AP
Sample : VIBLK
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 03/23/2022
Supervised By :Mahesh Dadoda 03/24/2022

Quant Time: Mar 23 03:41:04 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL032122AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Mar
QLast Update : Mon Mar 21 17:09:26 2022
Response via : Initial Calibration

