

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032222\
 Data File : VL038711.D
 Acq On : 23 Mar 2022 9:51
 Operator : SY/AP
 Sample : N1961-14DL2 20X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 C0P77DL2

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 03/24/2022
 Supervised By :Mahesh Dadoda 03/25/2022

Quant Time: Mar 24 02:54:34 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL032122AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Mar 2
 QLast Update : Mon Mar 21 17:09:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.648	49	671545	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	7.153	114	1783972	10.000	ppbv	-0.01
55) Chlorobenzene-d5	12.057	117	1542921	10.000	ppbv	-0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.388	95	1142806	9.661	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.600%
Target Compounds						
						Qvalue
3) Chlorodifluoromethane	2.980	51	10319	0.145	ppbv	95
13) Ethanol	3.613	45	8070m	2.136	ppbv	
20) Methylene Chloride	4.330	84	62753	2.571	ppbv	96
26) Hexane	5.671	57	44639	0.684	ppbv #	54
30) Cyclohexane	7.115	84	26749m	0.471	ppbv	
53) Toluene	9.780	91	1142149	7.720	ppbv	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032222\
Data File : VL038711.D
Acq On : 23 Mar 2022 9:51
Operator : SY/AP
Sample : N1961-14DL2 20X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
C0P77DL2

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 03/24/2022
Supervised By :Mahesh Dadoda 03/25/2022

Quant Time: Mar 24 02:54:34 2022

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL032122AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Mar

QLast Update : Mon Mar 21 17:09:26 2022

Response via : Initial Calibration

