

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL032320\
 Data File : VL034927.D
 Acq On : 23 Mar 2020 11:48
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Mar 24 02:01:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL030920AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Mar 09 16:42:53 2020
 QLast Update : Mon Mar 09 16:42:53 2020
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	105	0.00
2 T	Dichlorodifluoromethane	1.369	1.486	-8.5	148	0.00
3	Chlorodifluoromethane	1.117	1.104	1.2	114	0.00
4	Chloromethane	0.631	0.629	0.3	113	0.00
5 T	Vinyl Chloride	0.607	0.617	-1.6	114	0.00
6 T	Bromomethane	0.353	0.371	-5.1	119	0.00
7	Chloroethane	0.219	0.232	-5.9	121	0.00
8 T	Dichlorotetrafluoroethane	1.422	1.431	-0.6	118	0.00
9 T	Propene	0.533	0.491	7.9	107	0.00
10 T	Heptane	1.313	1.355	-3.2	114	0.00
11 T	Trichlorofluoromethane	1.457	1.446	0.8	117	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.020	1.042	-2.2	119	0.00
13	Ethanol	0.162	0.131	19.1	112	0.00
14 T	Bromoethene	0.440	0.444	-0.9	113	0.00
15 T	Acetone	1.296	1.157	10.7	117	0.00
16 T	1,3-Butadiene	0.586	0.580	1.0	113	0.00
17	tert-Butyl alcohol	1.249	0.974	22.0	94	0.00
18 T	1,1-Dichloroethene	0.452	0.479	-6.0	120	0.00
19 T	Isopropyl Alcohol	0.818	0.713	12.8	105	0.00
20 T	Methylene Chloride	0.410	0.423	-3.2	122	0.00
21 T	Allyl Chloride	0.832	0.817	1.8	112	0.00
22 T	trans-1,2-Dichloroethene	0.458	0.470	-2.6	117	0.00
23 T	Vinyl Acetate	1.915	1.809	5.5	106	0.00
24 T	1,1-Dichloroethane	1.283	1.262	1.6	112	0.00
25 T	Ethyl Acetate	2.309	2.219	3.9	111	0.00
26 T	Hexane	0.994	1.010	-1.6	115	0.00
27 T	Carbon Disulfide	1.128	1.249	-10.7	117	0.00
28 T	Methyl tert-Butyl Ether	1.654	1.507	8.9	102	0.00
29 T	Chloroform	1.517	1.536	-1.3	118	0.00
30 T	Cyclohexane	0.830	0.828	0.2	110	0.00
31 T	cis-1,2-Dichloroethene	1.017	1.006	1.1	111	0.00
32 T	1,1,1-Trichloroethane	1.478	1.493	-1.0	112	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	100	0.00
34 T	2-Butanone	0.665	0.666	-0.2	111	0.00
35 T	Carbon Tetrachloride	0.622	0.691	-11.1	111	0.00
36 T	Benzene	0.891	0.904	-1.5	110	0.00
37 T	1,2-Dichloroethane	0.503	0.539	-7.2	115	0.00
38 T	Trichloroethene	0.339	0.338	0.3	108	0.00
39 T	1,2-Dichloropropane	0.353	0.359	-1.7	112	0.00
40 T	1,4-Dioxane	0.133	0.128	3.8	112	0.00
41 T	Tetrahydrofuran	0.360	0.374	-3.9	108	0.00
42 T	Bromodichloromethane	0.685	0.756	-10.4	115	0.00
43	Methyl Methacrylate	0.376	0.407	-8.2	112	0.00
44 T	2,2,4-Trimethylpentane	1.506	1.569	-4.2	113	0.00
45 T	t-1,3-Dichloropropene	0.441	0.465	-5.4	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL032320\
 Data File : VL034927.D
 Acq On : 23 Mar 2020 11:48
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Mar 24 02:01:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL030920AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Mar 09 16:42:53 2020
 QLast Update : Mon Mar 09 16:42:53 2020
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	cis-1,3-Dichloropropene	0.524	0.551	-5.2	106	0.00
47 T	1,1,2-Trichloroethane	0.362	0.375	-3.6	111	0.00
48 T	Dibromochloromethane	0.587	0.625	-6.5	107	0.00
49 T	Bromoform	0.523	0.565	-8.0	105	0.00
50 T	4-Methyl-2-Pentanone	0.957	1.011	-5.6	112	0.00
51 T	2-Hexanone	0.777	0.827	-6.4	110	0.00
52 T	Tetrachloroethene	0.353	0.314	11.0	101	0.00
53 T	Toluene	1.029	1.051	-2.1	109	0.00
54 T	1,2-Dibromoethane	0.546	0.570	-4.4	110	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	104	0.00
56	1,1,1,2-Tetrachloroethane	0.449	0.431	4.0	107	0.00
57 T	Chlorobenzene	0.782	0.720	7.9	109	0.00
58 T	Ethyl Benzene	1.354	1.308	3.4	110	0.00
59 T	m/p-Xylene	1.135	1.072	5.6	111	0.00
60 T	o-Xylene	1.108	1.077	2.8	114	0.00
61 T	Styrene	0.777	0.773	0.5	108	0.00
62	Isopropylbenzene	1.675	1.590	5.1	111	0.00
63 T	1,1,2,2-Tetrachloroethane	0.752	0.769	-2.3	117	0.00
64	n-propylbenzene	0.436	0.412	5.5	107	0.00
65	tert-Butylbenzene	1.464	1.373	6.2	110	0.00
66 T	Benzyl Chloride	0.611	0.548	10.3	91	0.00
67	sec-Butylbenzene	2.030	1.969	3.0	111	0.00
68 S	1-Bromo-4-Fluorobenzene	0.780	0.784	-0.5	106	0.00
69	p-Isopropyltoluene	1.702	1.638	3.8	108	0.00
70	n-Butylbenzene	1.737	1.750	-0.7	113	0.00
71	2-Chlorotoluene	1.268	1.233	2.8	113	0.00
72 T	4-Ethyltoluene	1.308	1.284	1.8	111	0.00
73 T	1,3,5-Trimethylbenzene	1.225	1.196	2.4	110	0.00
74 T	1,2,4-Trimethylbenzene	1.262	1.248	1.1	112	0.00
75 T	1,3-Dichlorobenzene	0.761	0.716	5.9	110	0.00
76 T	1,4-Dichlorobenzene	0.759	0.718	5.4	109	0.00
77 T	1,2-Dichlorobenzene	0.737	0.701	4.9	108	0.00
78 T	Hexachloro-1,3-Butadiene	0.616	0.564	8.4	104	0.00
79 T	Naphthalene	0.984	1.174	-19.3	105	0.00
80 T	Naphthalene,2-methyl-	0.257	0.464	-80.5#	93	0.00
81 T	1,2,4-Trichlorobenzene	0.625	0.669	-7.0	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0