

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042176.D
 Acq On : 27 Mar 2025 10:32
 Operator : SY/MD
 Sample : VSTDICC0.1
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.1

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 03/31/2025
 Supervised By :Mahesh Dadoda 04/03/2025

Quant Time: Mar 28 01:35:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri Mar 28 01:31:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	180740	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.955	114	528028	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.878	117	481506	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.374	95	398069	10.067	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.700%
Target Compounds						
5) Vinyl Chloride	1.580	62	1274	0.115	ppbv #	42
32) 1,1,1-Trichloroethane	3.366	97	4509m	0.106	ppbv	
35) Carbon Tetrachloride	3.768	117	4131m	0.102	ppbv	
38) Trichloroethene	4.551	130	2306m	0.106	ppbv	
52) Tetrachloroethene	8.225	164	1895m	0.091	ppbv	
54) 1,2-Dibromoethane	7.652	107	3033m	0.108	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.966	83	3852m	0.093	ppbv	
79) Naphthalene	13.513	128	7882	0.084	ppbv #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
Data File : VL042176.D
Acq On : 27 Mar 2025 10:32
Operator : SY/MD
Sample : VSTDICC0.1
Misc : 400mL/MSVOA_L
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 03/31/2025
Supervised By :Mahesh Dadoda 04/03/2025

Quant Time: Mar 28 01:35:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Fri Mar 28 01:31:23 2025
Response via : Initial Calibration

