

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042181.D
 Acq On : 27 Mar 2025 14:39
 Operator : SY/MD
 Sample : VL0327ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0327ABS01

Quant Time: Mar 28 02:12:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri Mar 28 02:00:52 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 03/31/2025
 Supervised By : Mahesh Dadoda 04/03/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	166105	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.955	114	483209	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.881	117	452250	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.377	95	368941	9.934	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.300%

Target Compounds	Qvalue					
2) Dichlorodifluoromethane	1.499	85	243063	9.911	ppbv	100
3) Chlorodifluoromethane	1.476	51	241481	9.638	ppbv	99
4) Chloromethane	1.531	50	96241	9.602	ppbv	99
5) Vinyl Chloride	1.580	62	91865	8.649	ppbv	94
6) Bromomethane	1.667	94	40361	8.886	ppbv	92
7) Chloroethane	1.703	64	37307	9.273	ppbv	100
8) Dichlorotetrafluoroethane	1.554	85	208981	10.006	ppbv	94
9) Propene	1.482	41	102862	8.957	ppbv	99
10) Heptane	4.978	43	290621	9.514	ppbv	100
11) Trichlorofluoromethane	1.877	101	232086	9.725	ppbv	95
12) 1,1,2-Trichlorotrifluo...	2.140	101	178792	9.115	ppbv	99
13) Ethanol	1.719	45	5124	4.720	ppbv	93
14) Bromoethene	1.780	108	69775	9.286	ppbv	98
15) Acetone	1.835	43	195717	8.425	ppbv	96
16) 1,3-Butadiene	1.609	39	98877	8.716	ppbv	100
17) tert-Butyl alcohol	2.039	59	266424	9.208	ppbv	98
18) 1,1-Dichloroethene	2.033	96	83339	9.088	ppbv	95
19) Isopropyl Alcohol	1.884	45	120986	9.230	ppbv	97
20) Methylene Chloride	2.062	84	69930	8.083	ppbv	96
21) Allyl Chloride	2.094	41	147810	9.022	ppbv	99
22) trans-1,2-Dichloroethene	2.340	96	96616	9.116	ppbv	99
23) Vinyl Acetate	2.463	43	94696	9.418	ppbv	96
24) 1,1-Dichloroethane	2.408	63	189793	9.661	ppbv	99
25) Ethyl Acetate	2.835	43	574442	9.528	ppbv	99
26) Hexane	2.826	57	224704	9.649	ppbv	98
27) Carbon Disulfide	2.149	76	230410	9.226	ppbv #	94
28) Methyl tert-Butyl Ether	2.441	73	149961	10.342	ppbv	95
29) Chloroform	2.848	83	336404	9.870	ppbv	96
30) Cyclohexane	3.855	84	194220	9.604	ppbv	98
31) cis-1,2-Dichloroethene	2.719	61	231151	9.777	ppbv	97
32) 1,1,1-Trichloroethane	3.366	97	359860	9.223	ppbv	99
34) 2-Butanone	2.557	43	320843	9.157	ppbv	100
35) Carbon Tetrachloride	3.761	117	361037	9.565	ppbv	99
36) Benzene	3.658	78	452718	9.450	ppbv	98
37) 1,2-Dichloroethane	3.217	62	277247	10.109	ppbv	98
38) Trichloroethene	4.548	130	187209	9.366	ppbv	98
39) 1,2-Dichloropropane	4.315	63	164374	9.422	ppbv	95
40) 1,4-Dioxane	4.583	88	82012	8.857	ppbv #	92
41) Tetrahydrofuran	3.052	42	188907	9.362	ppbv	99
42) Bromodichloromethane	4.489	83	377092	9.974	ppbv	98
43) Methyl Methacrylate	4.881	69	190156	9.377	ppbv	100
44) 2,2,4-Trimethylpentane	4.638	57	769803	9.509	ppbv	99
45) t-1,3-Dichloropropene	6.415	75	204004	9.376	ppbv	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042181.D
 Acq On : 27 Mar 2025 14:39
 Operator : SY/MD
 Sample : VL0327ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0327ABS01

Manual Integrations
 APPROVED

Quant Time: Mar 28 02:12:34 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Mar 28 02:00:52 2025

Response via : Initial Calibration

Reviewed By : Semsettin Yesilyurt 03/31/2025
 Supervised By : Mahesh Dadoda 04/03/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.599	75	251694	9.437	ppbv	97
47) 1,1,2-Trichloroethane	6.583	97	175509	9.517	ppbv	99
48) Dibromochloromethane	7.386	129	310158	9.923	ppbv	99
49) Bromoform	9.483	173	279382	9.957	ppbv	99
50) 4-Methyl-2-Pentanone	5.748	43	481919	8.679	ppbv	98
51) 2-Hexanone	7.435	43	395520	8.674	ppbv	99
52) Tetrachloroethene	8.224	164	167015	8.791	ppbv	95
53) Toluene	6.930	91	540612	9.520	ppbv	99
54) 1,2-Dibromoethane	7.655	107	246519	9.589	ppbv	99
56) 1,1,1,2-Tetrachloroethane	8.924	131	240905	9.677	ppbv	96
57) Chlorobenzene	8.920	112	407816	9.507	ppbv	98
58) Ethyl Benzene	9.354	91	769818	9.583	ppbv	97
59) m/p-Xylene	9.551	91	1220553m	19.028	ppbv	
60) o-Xylene	9.962	91	612011	9.518	ppbv	98
61) Styrene	9.869	104	306658	9.639	ppbv	99
62) Isopropylbenzene	10.539	105	923163	9.600	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.962	83	370486	9.483	ppbv	99
64) n-propylbenzene	10.989	120	239652	9.682	ppbv	98
65) tert-Butylbenzene	11.532	119	828438	9.315	ppbv	98
66) Benzyl Chloride	11.616	91	81404	8.627	ppbv	98
67) sec-Butylbenzene	11.769	105	1155378	9.347	ppbv	99
69) p-Isopropyltoluene	11.927	119	1011573	9.503	ppbv	99
70) n-Butylbenzene	12.283	91	1022935	9.397	ppbv	99
71) 2-Chlorotoluene	10.901	91	713648	9.562	ppbv	98
72) 4-Ethyltoluene	11.128	105	766720	9.576	ppbv	100
73) 1,3,5-Trimethylbenzene	11.205	105	660905	9.451	ppbv	99
74) 1,2,4-Trimethylbenzene	11.539	105	718683	8.921	ppbv	97
75) 1,3-Dichlorobenzene	11.610	146	415378	9.401	ppbv	100
76) 1,4-Dichlorobenzene	11.671	146	418670	9.521	ppbv	99
77) 1,2-Dichlorobenzene	11.947	146	401634	9.336	ppbv	99
78) Hexachloro-1,3-Butadiene	13.882	225	346566	7.914	ppbv	100
79) Naphthalene	13.513	128	762976	8.659	ppbv	99
80) Naphthalene,2-methyl-	14.458	142	308264	7.134	ppbv	98
81) 1,2,4-Trichlorobenzene	13.442	180	356193	8.764	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
Data File : VL042181.D
Acq On : 27 Mar 2025 14:39
Operator : SY/MD
Sample : VL0327ABS01
Misc : 400mL/MSVOA_L
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0327ABS01

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 03/31/2025
Supervised By : Mahesh Dadoda 04/03/2025

Quant Time: Mar 28 02:12:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Fri Mar 28 02:00:52 2025
Response via : Initial Calibration

