

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL040419\
Data File : VL033532.D
Acq On : 4 Apr 2019 14:10
Operator : SY/AP
Sample : VL0404ABL01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0404ABL01

Quant Time: Apr 05 02:18:19 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Mar 26 16:00:50 2019
QLast Update : Tue Mar 26 16:00:50 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	1177523	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.28	114	2671743	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.24	117	2703961	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.57	95	2132145	10.06	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.60%

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL040419\
Data File : VL033532.D
Acq On : 4 Apr 2019 14:10
Operator : SY/AP
Sample : VL0404ABL01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0404ABL01

Quant Time: Apr 05 02:18:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Mar 26 16:00:50 2019
QLast Update : Tue Mar 26 16:00:50 2019
Response via : Initial Calibration

