

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL040419\
Data File : VL033534.D
Acq On : 4 Apr 2019 15:43
Operator : SY/AP
Sample : QC040319 CAN 10284
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC040319 CAN 10284

Quant Time: Apr 05 02:21:08 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Mar 26 16:00:50 2019
QLast Update : Tue Mar 26 16:00:50 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	1242378	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.28	114	2815620	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.23	117	2968859	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.57	95	2328267	10.00	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.00%

Target Compounds					Ovalue
20) Methylene Chloride	4.41	84	18701	0.295 ppbv	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL040419\
Data File : VL033534.D
Acq On : 4 Apr 2019 15:43
Operator : SY/AP
Sample : QC040319 CAN 10284
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC040319 CAN 10284

Quant Time: Apr 05 02:21:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LTue Mar 26 16:00:50 2019
QLast Update : Tue Mar 26 16:00:50 2019
Response via : Initial Calibration

