

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL040424\
 Data File : VL041128.D
 Acq On : 4 Apr 2024 16:48
 Operator : SY/MD
 Sample : VL0404ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0404ABL01

Quant Time: Apr 05 05:10:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL040424AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Apr 05 04:49:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.214	49	1386951	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.677	114	3501309	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.500	117	3050748	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.809	95	2426170	9.890	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL040424\
Data File : VL041128.D
Acq On : 4 Apr 2024 16:48
Operator : SY/MD
Sample : VL0404ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0404ABL01

Quant Time: Apr 05 05:10:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL040424AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri Apr 05 04:49:46 2024
Response via : Initial Calibration

