

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL040424\
 Data File : VL041130.D
 Acq On : 4 Apr 2024 18:06
 Operator : SY/MD
 Sample : QC040224 CAN 10606
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC040224 CAN 10606

Quant Time: Apr 05 05:11:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL040424AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Apr 05 04:49:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.211	49	1394462	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.677	114	3501291	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.497	117	3038833	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.806	95	2422543	9.914	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL040424\
Data File : VL041130.D
Acq On : 4 Apr 2024 18:06
Operator : SY/MD
Sample : QC040224 CAN 10606
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC040224 CAN 10606

Quant Time: Apr 05 05:11:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL040424AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri Apr 05 04:49:46 2024
Response via : Initial Calibration

