

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL040424\
 Data File : VL041138.D
 Acq On : 4 Apr 2024 23:31
 Operator : SY/MD
 Sample : P1953-04DL 20X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV4DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/05/2024
 Supervised By :Mahesh Dadoda 04/05/2024

Quant Time: Apr 05 05:14:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL040424AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LFri Aug 2
 QLast Update : Fri Apr 05 04:49:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.211	49	1385406	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.674	114	3893544	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.491	117	3507298	10.000	ppbv	-0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.799	95	2684341	9.518	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	95.200%
Target Compounds						
					Qvalue	
13) Ethanol	3.269	45	24620m	1.525	ppbv	
15) Acetone	3.494	43	179470	1.196	ppbv #	79
20) Methylene Chloride	3.947	84	73100	1.182	ppbv	97
26) Hexane	5.234	57	30298m	0.240	ppbv	
34) 2-Butanone	4.816	43	436404m	2.649	ppbv	
51) 2-Hexanone	9.600	43	50346m	0.250	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL040424\
Data File : VL041138.D
Acq On : 4 Apr 2024 23:31
Operator : SY/MD
Sample : P1953-04DL 20X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV4DL

Quant Time: Apr 05 05:14:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL040424AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Apr 05 04:49:46 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 04/05/2024
Supervised By : Mahesh Dadoda 04/05/2024

