

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL040424\
 Data File : VL041142.D
 Acq On : 5 Apr 2024 2:16
 Operator : SY/MD
 Sample : P1986-01DL 20X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV1DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/05/2024
 Supervised By :Mahesh Dadoda 04/05/2024

Quant Time: Apr 05 05:16:27 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL040424AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Fri Apr 05 04:49:46 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.211	49	1338329	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.674	114	3772156	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.491	117	3348079	10.000	ppbv	-0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.799	95	2586311	9.606	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.100%
Target Compounds						
					Qvalue	
15) Acetone	3.501	43	66872	0.461	ppbv #	63
20) Methylene Chloride	3.948	84	10043	0.168	ppbv #	83
26) Hexane	5.224	57	12698m	0.104	ppbv	
34) 2-Butanone	4.819	43	121797	0.763	ppbv	97
51) 2-Hexanone	9.597	43	36176m	0.186	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL040424\
Data File : VL041142.D
Acq On : 5 Apr 2024 2:16
Operator : SY/MD
Sample : P1986-01DL 20X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV1DL

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 04/05/2024
Supervised By : Mahesh Dadoda 04/05/2024

Quant Time: Apr 05 05:16:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL040424AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Apr 05 04:49:46 2024
Response via : Initial Calibration

