

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL040424\
 Data File : VL041125.D
 Acq On : 4 Apr 2024 12:40
 Operator : SY/MD
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/05/2024
 Supervised By :Mahesh Dadoda 04/05/2024

Quant Time: Apr 05 04:18:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL040424AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LFri Aug 2
 QLast Update : Fri Apr 05 04:13:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.214	49	1366828	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.681	114	3410305	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.500	117	3013728	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.812	95	2446568	10.132	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.300%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	2.932	62	2019	0.027	ppbv	94
32) 1,1,1-Trichloroethane	6.025	97	5386m	0.028	ppbv	
35) Carbon Tetrachloride	6.513	117	4828m	0.025	ppbv	
38) Trichloroethene	7.317	130	2665m	0.029	ppbv	
52) Tetrachloroethene	10.648	164	2331m	0.029	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.092	83	6496m	0.029	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL040424\
Data File : VL041125.D
Acq On : 4 Apr 2024 12:40
Operator : SY/MD
Sample : VSTDICC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Quant Time: Apr 05 04:18:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL040424AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Apr 05 04:13:19 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 04/05/2024
Supervised By :Mahesh Dadoda 04/05/2024

