

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL040621\
 Data File : VL036704.D
 Acq On : 6 Apr 2021 19:10
 Operator : SY/AP
 Sample : M1929-05DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 FS-OA-04DL

Manual Integrations
 APPROVED

MMDadoda
 4/12/2021 11:10:59 AM

Quant Time: Apr 08 12:25:45 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar 25 2021
 QLast Update : Thu Mar 25 09:42:40 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1266384	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	3045812	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.10	117	2673362	10.00	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.43	95	2044397	9.59	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	95.90%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.62	45	36578m	4.072	ppbv	
15) Acetone	3.86	43	142583m	1.144	ppbv	
19) Isopropyl Alcohol	4.00	45	35358m	0.516	ppbv	
20) Methylene Chloride	4.35	84	132743m	2.275	ppbv	
26) Hexane	5.69	57	98848	0.719	ppbv #	45
53) Toluene	9.82	91	30661	0.102	ppbv	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL040621\
Data File : VL036704.D
Acq On : 6 Apr 2021 19:10
Operator : SY/AP
Sample : M1929-05DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
FS-OA-04DL

Manual Integrations
APPROVED

MMDadoda
4/12/2021 11:10:59 AM

Quant Time: Apr 08 12:25:45 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar
QLast Update : Thu Mar 25 09:42:40 2021
Response via : Initial Calibration

