

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL040621\
 Data File : VL036714.D
 Acq On : 7 Apr 2021 1:44
 Operator : SY/AP
 Sample : M1929-01 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 FS-SV1-FIELD

Manual Integrations
 APPROVED

MMDadoda
 4/12/2021 11:11:37 AM

Quant Time: Apr 07 06:32:06 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar 25 2021
 QLast Update : Thu Mar 25 09:42:40 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1141378	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.17	114	2989567	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.09	117	2584255	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	1971100	9.57	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	95.70%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.02	41	12339m	0.181	ppbv	
13) Ethanol	3.62	45	42148m	5.206	ppbv	
15) Acetone	3.86	43	128827m	1.147	ppbv	
19) Isopropyl Alcohol	4.00	45	23259m	0.376	ppbv	
20) Methylene Chloride	4.34	84	67028	1.274	ppbv	91
26) Hexane	5.68	57	38829	0.313	ppbv #	43

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL040621\
Data File : VL036714.D
Acq On : 7 Apr 2021 1:44
Operator : SY/AP
Sample : M1929-01 10X
Misc : 400mL/MSVOA L
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
FS-SV1-FIELD

Manual Integrations
APPROVED

MMDadoda
4/12/2021 11:11:37 AM

Quant Time: Apr 07 06:32:06 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar
QLast Update : Thu Mar 25 09:42:40 2021
Response via : Initial Calibration

