

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL040722\
 Data File : VL038837.D
 Acq On : 7 Apr 2022 12:20
 Operator : SY/AP
 Sample : QC040622 CAN 10271
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC040622 CAN 10271

Quant Time: Apr 08 05:24:11 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL040522AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Apr 06 06:12:53 2022
 QLast Update : Wed Apr 06 06:12:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.680	49	1267036	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	7.185	114	3379870	10.000	ppbv	0.01
55) Chlorobenzene-d5	12.092	117	2906518	10.000	ppbv	0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.423	95	2250001	10.206	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL040722\
Data File : VL038837.D
Acq On : 7 Apr 2022 12:20
Operator : SY/AP
Sample : QC040622 CAN 10271
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC040622 CAN 10271

Quant Time: Apr 08 05:24:11 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL040522AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Apr 06 06:12:53 2022
QLast Update : Wed Apr 06 06:12:53 2022
Response via : Initial Calibration

