

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL040722\
 Data File : VL038838.D
 Acq On : 7 Apr 2022 12:57
 Operator : SY/AP
 Sample : QC040622 CAN 10271
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC040622 CAN 10271

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/09/2022
 Supervised By :Mahesh Dadoda 04/11/2022

Quant Time: Apr 08 05:24:18 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL040522AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Apr 06 06:12:53 2022
 QLast Update : Wed Apr 06 06:12:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.681	49	1238906	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	7.189	114	3312687	10.000	ppbv	0.02
55) Chlorobenzene-d5	12.098	117	2930465	10.000	ppbv	0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.429	95	2288417	10.295	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.900%
Target Compounds						
20) Methylene Chloride	4.353	84	7430m	0.166	ppbv	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL040722\
Data File : VL038838.D
Acq On : 7 Apr 2022 12:57
Operator : SY/AP
Sample : QC040622 CAN 10271
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC040622 CAN 10271

Quant Time: Apr 08 05:24:18 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL040522AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Apr
QLast Update : Wed Apr 06 06:12:53 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 04/09/2022
Supervised By :Mahesh Dadoda 04/11/2022

