

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL040722\
 Data File : VL038841.D
 Acq On : 7 Apr 2022 14:56
 Operator : SY/AP
 Sample : N2245-01DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BD-INDOORDL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/09/2022
 Supervised By :Mahesh Dadoda 04/11/2022

Quant Time: Apr 08 05:25:01 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL040522AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LWed Apr 0
 QLast Update : Wed Apr 06 06:12:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.684	49	1199093	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	7.192	114	3185251	10.000	ppbv	0.02
55) Chlorobenzene-d5	12.101	117	2760598	10.000	ppbv	0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.433	95	2078835	9.928	ppbv	0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.300%
Target Compounds						
					Qvalue	
13) Ethanol	3.620	45	51373m	10.918	ppbv	
20) Methylene Chloride	4.362	84	293285	6.774	ppbv	94
26) Hexane	5.706	57	278322	2.019	ppbv #	40

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL040722\
Data File : VL038841.D
Acq On : 7 Apr 2022 14:56
Operator : SY/AP
Sample : N2245-01DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BD-INDOORDL

Quant Time: Apr 08 05:25:01 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL040522AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Apr
QLast Update : Wed Apr 06 06:12:53 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 04/09/2022
Supervised By :Mahesh Dadoda 04/11/2022

