

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL040722\
 Data File : VL038843.D
 Acq On : 7 Apr 2022 16:15
 Operator : SY/AP
 Sample : N2245-02DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BD-AMBIENTDL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/09/2022
 Supervised By :Mahesh Dadoda 04/11/2022

Quant Time: Apr 08 05:25:28 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL040522AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Apr 06 06:12:53 2022
 QLast Update : Wed Apr 06 06:12:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.687	49	1251925	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	7.195	114	3116882	10.000	ppbv	0.02
55) Chlorobenzene-d5	12.105	117	2690428	10.000	ppbv	0.03
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.433	95	2079036	10.187	ppbv	0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.900%
Target Compounds						
					Qvalue	
9) Propene	3.038	41	11703m	0.145	ppbv	
13) Ethanol	3.639	45	28455m	5.792	ppbv	
15) Acetone	3.886	43	58316	0.468	ppbv #	87
20) Methylene Chloride	4.369	84	53596	1.186	ppbv	92
26) Hexane	5.706	57	61815	0.429	ppbv #	40
53) Toluene	9.825	91	100512	0.307	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL040722\
Data File : VL038843.D
Acq On : 7 Apr 2022 16:15
Operator : SY/AP
Sample : N2245-02DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BD-AMBIENTDL

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 04/09/2022
Supervised By : Mahesh Dadoda 04/11/2022

Quant Time: Apr 08 05:25:28 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL040522AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Apr
QLast Update : Wed Apr 06 06:12:53 2022
Response via : Initial Calibration

