

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL040722\
 Data File : VL038862.D
 Acq On : 8 Apr 2022 4:38
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/09/2022
 Supervised By :Mahesh Dadoda 04/11/2022

Quant Time: Apr 08 10:01:02 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL040522A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Wed Apr 06 06:12:53 2022
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.70	49	1041254	10.00	ppbv	0.04
33) 1,4-Difluorobenzene	7.20	114	2604655	10.00	ppbv	0.03
55) Chlorobenzene-d5	12.11	117	2336595	10.00	ppbv	0.04

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.44	95	1833199	10.34	ppbv	0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.40%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
20) Methylene Chloride	4.37	84	5848m	0.156	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL040722\
Data File : VL038862.D
Acq On : 8 Apr 2022 4:38
Operator : SY/AP
Sample : VIBLK
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

Quant Time: Apr 08 10:01:02 2022
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL040522
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Wed Apr 06 06:12:53 2022
Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 04/09/2022
Supervised By :Mahesh Dadoda 04/11/2022

