

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL040725\
 Data File : VL042238.D
 Acq On : 07 Apr 2025 18:46
 Operator : SY/MD
 Sample : Q1725-03DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SSSG-DUP-2-04012025DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/08/2025
 Supervised By :Mahesh Dadoda 04/09/2025

Quant Time: Apr 08 01:34:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LFri Aug 2
 QLast Update : Fri Mar 28 02:00:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.803	49	200446	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	3.985	114	518231	10.000	ppbv	0.03
55) Chlorobenzene-d5	8.904	117	450849	10.000	ppbv	0.03
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.396	95	356358	9.625	ppbv	0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.200%
Target Compounds						
13) Ethanol	1.732	45	2273m	1.735	ppbv	Qvalue
15) Acetone	1.845	43	119931	4.278	ppbv	95
19) Isopropyl Alcohol	1.897	45	221838	14.025	ppbv #	32
20) Methylene Chloride	2.075	84	2236	0.214	ppbv	92
59) m/p-Xylene	9.555	91	23324m	0.365	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL040725\
Data File : VL042238.D
Acq On : 07 Apr 2025 18:46
Operator : SY/MD
Sample : Q1725-03DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SSSG-DUP-2-04012025DL

Quant Time: Apr 08 01:34:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Mar 28 02:00:52 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 04/08/2025
Supervised By :Mahesh Dadoda 04/09/2025

