

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL040824\
 Data File : VL041148.D
 Acq On : 8 Apr 2024 13:21
 Operator : SY/MD
 Sample : P1953-03DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV3DL

Quant Time: Apr 09 01:13:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL040424AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Apr 09 01:10:08 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.182	49	1270060	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.645	114	3447505	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.458	117	3102809	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.767	95	2442528	9.790	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.900%
Target Compounds						
					Qvalue	
9) Propene	2.694	41	16493	0.265	ppbv	99
13) Ethanol	3.253	45	13249	0.895	ppbv	84
15) Acetone	3.465	43	382857	2.784	ppbv #	75
20) Methylene Chloride	3.928	84	170272	3.002	ppbv	97
26) Hexane	5.202	57	94182	0.813	ppbv #	39
34) 2-Butanone	4.784	43	957251	6.562	ppbv	99
51) 2-Hexanone	9.555	43	109383	0.615	ppbv #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL040824\
Data File : VL041148.D
Acq On : 8 Apr 2024 13:21
Operator : SY/MD
Sample : P1953-03DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV3DL

Quant Time: Apr 09 01:13:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL040424AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Tue Apr 09 01:10:08 2024
Response via : Initial Calibration

