

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL040921\
 Data File : VL036781.D
 Acq On : 9 Apr 2021 19:25
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Manual Integrations
 APPROVED

MMDadoda
 4/12/2021 5:42:56 PM

Quant Time: Apr 10 07:30:34 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar 25 2021
 QLast Update : Thu Mar 25 09:42:40 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1241189	10.00	ppbv	-0.01
33) 1,4-Difluorobenzene	7.17	114	3073267	10.00	ppbv	-0.01
55) Chlorobenzene-d5	12.08	117	2393051	10.00	ppbv	-0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.41	95	1571004	8.24	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	82.40%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	3.87	43	25247m	0.207	ppbv	
20) Methylene Chloride	4.34	84	7035	0.123	ppbv #	62

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL040921\
Data File : VL036781.D
Acq On : 9 Apr 2021 19:25
Operator : SY/AP
Sample : VIBLK
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

MMDadoda
4/12/2021 5:42:56 PM

Quant Time: Apr 10 07:30:34 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar
QLast Update : Thu Mar 25 09:42:40 2021
Response via : Initial Calibration

