

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041125\
 Data File : VL042312.D
 Acq On : 11 Apr 2025 09:50
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 04/16/2025
 Supervised By : Mahesh Dadoda 04/17/2025

Quant Time: Apr 11 22:56:21 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2025

QLast Update : Fri Mar 28 02:00:52 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	183700	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.959	114	463746	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.885	117	401340	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.377 95 318204 9.655 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 96.500%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	264775	9.762	ppbv	100
3) Chlorodifluoromethane	1.476	51	287991	10.394	ppbv	97
4) Chloromethane	1.531	50	111786	10.085	ppbv	98
5) Vinyl Chloride	1.580	62	102342	8.713	ppbv	95
6) Bromomethane	1.667	94	43054	8.571	ppbv	99
7) Chloroethane	1.703	64	41946	9.427	ppbv	98
8) Dichlorotetrafluoroethane	1.554	85	224489	9.719	ppbv	94
9) Propene	1.483	41	117677	9.266	ppbv	99
10) Heptane	4.981	43	334707	9.908	ppbv	100
11) Trichlorofluoromethane	1.877	101	250122	9.477	ppbv	99
12) 1,1,2-Trichlorotrifluoro...	2.140	101	193510	8.920	ppbv	97
13) Ethanol	1.722	45	5849	4.872	ppbv	89
14) Bromoethene	1.780	108	73294	8.820	ppbv	100
15) Acetone	1.835	43	231724	9.019	ppbv	97
16) 1,3-Butadiene	1.609	39	112297	8.950	ppbv	97
17) tert-Butyl alcohol	2.039	59	255575	7.987	ppbv	99
18) 1,1-Dichloroethene	2.033	96	89512	8.827	ppbv	98
19) Isopropyl Alcohol	1.887	45	125233	8.639	ppbv	92
20) Methylene Chloride	2.062	84	76968	8.044	ppbv	93
21) Allyl Chloride	2.098	41	162998	8.996	ppbv	97
22) trans-1,2-Dichloroethene	2.340	96	101974	8.700	ppbv	95
23) Vinyl Acetate	2.463	43	95152	8.557	ppbv	99
24) 1,1-Dichloroethane	2.408	63	207862	9.567	ppbv	99
25) Ethyl Acetate	2.835	43	678920	10.182	ppbv	99
26) Hexane	2.826	57	256866	9.973	ppbv	98
27) Carbon Disulfide	2.149	76	243367	8.812	ppbv	98
28) Methyl tert-Butyl Ether	2.441	73	129141	8.053	ppbv	96
29) Chloroform	2.848	83	393638	10.443	ppbv	98
30) Cyclohexane	3.858	84	203526	9.100	ppbv	97
31) cis-1,2-Dichloroethene	2.722	61	247194	9.454	ppbv	97
32) 1,1,1-Trichloroethane	3.370	97	380312	8.813	ppbv	98
34) 2-Butanone	2.557	43	380682	11.321	ppbv	98
35) Carbon Tetrachloride	3.764	117	390140	10.769	ppbv	98
36) Benzene	3.658	78	506278	11.012	ppbv	97
37) 1,2-Dichloroethane	3.221	62	292183	11.101	ppbv	99
38) Trichloroethene	4.551	130	203156	10.591	ppbv	96
39) 1,2-Dichloropropane	4.315	63	185497	11.079	ppbv	97
40) 1,4-Dioxane	4.583	88	82376	9.270	ppbv #	98
41) Tetrahydrofuran	3.056	42	216646	11.188	ppbv	93
42) Bromodichloromethane	4.493	83	413250	11.389	ppbv	99
43) Methyl Methacrylate	4.884	69	184472	9.478	ppbv	96
44) 2,2,4-Trimethylpentane	4.642	57	872871	11.235	ppbv	98
45) t-1,3-Dichloropropene	6.422	75	176804	8.467	ppbv	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041125\
 Data File : VL042312.D
 Acq On : 11 Apr 2025 09:50
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/16/2025
 Supervised By :Mahesh Dadoda 04/17/2025

Quant Time: Apr 11 22:56:21 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Mar 28 02:00:52 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.603	75	236039	9.222	ppbv	99
47) 1,1,2-Trichloroethane	6.583	97	202891	11.463	ppbv	98
48) Dibromochloromethane	7.389	129	333128	11.106	ppbv	99
49) Bromoform	9.484	173	294489	10.935	ppbv	100
50) 4-Methyl-2-Pentanone	5.755	43	531851	9.980	ppbv	95
51) 2-Hexanone	7.438	43	424035	9.689	ppbv #	94
52) Tetrachloroethene	8.224	164	170989	9.378	ppbv	97
53) Toluene	6.936	91	547579	10.048	ppbv	99
54) 1,2-Dibromoethane	7.655	107	267225	10.831	ppbv	99
56) 1,1,1,2-Tetrachloroethane	8.927	131	271350	12.283	ppbv	100
57) Chlorobenzene	8.924	112	460598	12.099	ppbv	98
58) Ethyl Benzene	9.357	91	779227	10.931	ppbv	98
59) m/p-Xylene	9.552	91	1323463m	23.249	ppbv	
60) o-Xylene	9.966	91	660065	11.567	ppbv	96
61) Styrene	9.872	104	291355	10.320	ppbv	99
62) Isopropylbenzene	10.542	105	988337	11.582	ppbv	98
63) 1,1,2,2-Tetrachloroethane	9.966	83	446341	12.873	ppbv	99
64) n-propylbenzene	10.992	120	268681	12.231	ppbv	97
65) tert-Butylbenzene	11.532	119	932850	11.820	ppbv	98
66) Benzyl Chloride	11.616	91	68675	8.201	ppbv	99
67) sec-Butylbenzene	11.769	105	1310042	11.942	ppbv	100
69) p-Isopropyltoluene	11.927	119	1107045	11.720	ppbv	99
70) n-Butylbenzene	12.283	91	1144571	11.848	ppbv	99
71) 2-Chlorotoluene	10.904	91	756562	11.423	ppbv	98
72) 4-Ethyltoluene	11.128	105	847821	11.932	ppbv	99
73) 1,3,5-Trimethylbenzene	11.205	105	729887	11.762	ppbv	99
74) 1,2,4-Trimethylbenzene	11.539	105	819490	11.463	ppbv	92
75) 1,3-Dichlorobenzene	11.610	146	492549	12.562	ppbv	99
76) 1,4-Dichlorobenzene	11.675	146	488636	12.522	ppbv	98
77) 1,2-Dichlorobenzene	11.950	146	494739	12.959	ppbv	98
78) Hexachloro-1,3-Butadiene	13.882	225	373025	9.599	ppbv	99
79) Naphthalene	13.516	128	824947	10.550	ppbv	100
80) Naphthalene,2-methyl-	14.458	142	276556	7.212	ppbv	98
81) 1,2,4-Trichlorobenzene	13.442	180	374714	10.389	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041125\
 Data File : VL042312.D
 Acq On : 11 Apr 2025 09:50
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 04/16/2025
 Supervised By : Mahesh Dadoda 04/17/2025

Quant Time: Apr 11 22:56:21 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug

QLast Update : Fri Mar 28 02:00:52 2025

Response via : Initial Calibration

