

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041222\
 Data File : VL038894.D
 Acq On : 12 Apr 2022 16:21
 Operator : SY/AP
 Sample : QC041122 CAN 10303
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC041122 CAN 10303

Quant Time: Apr 13 01:18:18 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL040522AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Apr 06 06:12:53 2022
 QLast Update : Wed Apr 06 06:12:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.664	49	1161821	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	7.169	114	2873478	10.000	ppbv	0.00
55) Chlorobenzene-d5	12.079	117	2501399	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.407	95	1857178	9.788	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041222\
Data File : VL038894.D
Acq On : 12 Apr 2022 16:21
Operator : SY/AP
Sample : QC041122 CAN 10303
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC041122 CAN 10303

Quant Time: Apr 13 01:18:18 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL040522AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Apr 06 06:12:53 2022
QLast Update : Wed Apr 06 06:12:53 2022
Response via : Initial Calibration

