

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041222\
 Data File : VL038895.D
 Acq On : 12 Apr 2022 17:08
 Operator : SY/AP
 Sample : N2339-01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SVE-EFFLUENT

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/13/2022
 Supervised By :Mahesh Dadoda 04/14/2022

Quant Time: Apr 13 01:18:26 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL040522AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Apr 06 06:12:53 2022
 QLast Update : Wed Apr 06 06:12:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.661	49	1048045	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	7.169	114	2667994	10.000	ppbv	0.00
55) Chlorobenzene-d5	12.079	117	2253859	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.404	95	1792008	10.482	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.800%
Target Compounds						
					Qvalue	
9) Propene	3.018	41	15066m	0.223	ppbv	
13) Ethanol	3.597	45	160889m	39.122	ppbv	
15) Acetone	3.854	43	56702	0.543	ppbv	100
20) Methylene Chloride	4.340	84	376856	9.959	ppbv	96
26) Hexane	5.681	57	530625	4.403	ppbv #	40
52) Tetrachloroethene	11.211	164	4398m	0.052	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041222\
Data File : VL038895.D
Acq On : 12 Apr 2022 17:08
Operator : SY/AP
Sample : N2339-01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SVE-EFFLUENT

Quant Time: Apr 13 01:18:26 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL040522AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Apr
QLast Update : Wed Apr 06 06:12:53 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 04/13/2022
Supervised By :Mahesh Dadoda 04/14/2022

