

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL041222\
 Data File : VL038889.D
 Acq On : 12 Apr 2022 12:54
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampled :
 VSTDCCC010

Quant Time: Apr 13 01:20:48 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL040522AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Apr 06 06:12:53 2022
 QLast Update : Wed Apr 06 06:12:53 2022
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	72	0.00
2 T	Dichlorodifluoromethane	1.418	1.237	12.8	70	0.00
3	Chlorodifluoromethane	1.220	1.177	3.5	74	0.00
4	Chloromethane	0.632	0.596	5.7	71	0.00
5 T	Vinyl Chloride	0.584	0.552	5.5	71	0.00
6 T	Bromomethane	0.283	0.276	2.5	74	0.00
7	Chloroethane	0.195	0.190	2.6	71	0.00
8 T	Dichlorotetrafluoroethane	1.325	1.323	0.2	78	0.00
9 T	Propene	0.646	0.617	4.5	72	0.00
10 T	Heptane	1.626	1.521	6.5	72	0.00
11 T	Trichlorofluoromethane	1.140	1.195	-4.8	79	0.00
12 T	1,1,2-Trichlorotrifluoroeth	0.873	0.897	-2.7	78	0.00
13	Ethanol	0.039	0.037#	5.1	78	0.00
14 T	Bromoethene	0.384	0.378	1.6	77	0.00
15 T	Acetone	0.996	0.932	6.4	76	0.00
16 T	1,3-Butadiene	0.569	0.550	3.3	73	0.00
17	tert-Butyl alcohol	0.893	0.730	18.3	58	0.00
18 T	1,1-Dichloroethene	0.398	0.400	-0.5	78	0.00
19 T	Isopropyl Alcohol	0.509	0.490	3.7	71	0.00
20 T	Methylene Chloride	0.361	0.326	9.7	71	0.00
21 T	Allyl Chloride	0.745	0.704	5.5	73	0.00
22 T	trans-1,2-Dichloroethene	0.403	0.434	-7.7	76	0.00
23 T	Vinyl Acetate	0.970	0.996	-2.7	76	0.00
24 T	1,1-Dichloroethane	0.876	0.871	0.6	73	0.00
25 T	Ethyl Acetate	2.526	2.472	2.1	72	0.00
26 T	Hexane	1.150	1.089	5.3	74	0.00
27 T	Carbon Disulfide	1.054	1.059	-0.5	72	0.00
28 T	Methyl tert-Butyl Ether	0.888	0.819	7.8	68	0.00
29 T	Chloroform	1.609	1.619	-0.6	77	0.00
30 T	Cyclohexane	1.013	0.940	7.2	73	0.00
31 T	cis-1,2-Dichloroethene	1.057	1.056	0.1	75	0.00
32 T	1,1,1-Trichloroethane	1.621	1.591	1.9	75	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	72	0.00
34 T	2-Butanone	0.395	0.393	0.5	73	0.00
35 T	Carbon Tetrachloride	0.600	0.600	0.0	75	0.00
36 T	Benzene	0.910	0.866	4.8	74	0.00
37 T	1,2-Dichloroethane	0.416	0.443	-6.5	78	0.00
38 T	Trichloroethene	0.373	0.339	9.1	72	0.00
39 T	1,2-Dichloropropane	0.348	0.334	4.0	73	0.00
40 T	1,4-Dioxane	0.118	0.113	4.2	81	0.00
41 T	Tetrahydrofuran	0.375	0.366	2.4	74	0.00
42 T	Bromodichloromethane	0.657	0.672	-2.3	77	0.00
43	Methyl Methacrylate	0.348	0.360	-3.4	74	0.00
44 T	2,2,4-Trimethylpentane	1.540	1.428	7.3	74	0.00
45 T	t-1,3-Dichloropropene	0.364	0.381	-4.7	69	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL041222\
 Data File : VL038889.D
 Acq On : 12 Apr 2022 12:54
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Apr 13 01:20:48 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL040522AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Apr 06 06:12:53 2022
 QLast Update : Wed Apr 06 06:12:53 2022
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	cis-1,3-Dichloropropene	0.480	0.483	-0.6	70	0.00
47 T	1,1,2-Trichloroethane	0.357	0.348	2.5	74	0.00
48 T	Dibromochloromethane	0.559	0.566	-1.3	73	0.00
49 T	Bromoform	0.443	0.442	0.2	72	0.00
50 T	4-Methyl-2-Pentanone	0.889	0.894	-0.6	73	0.00
51 T	2-Hexanone	0.670	0.718	-7.2	75	0.00
52 T	Tetrachloroethene	0.318	0.284	10.7	71	0.00
53 T	Toluene	1.050	0.999	4.9	74	0.00
54 T	1,2-Dibromoethane	0.495	0.502	-1.4	75	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	77	0.00
56	1,1,1,2-Tetrachloroethane	0.438	0.403	8.0	74	0.00
57 T	Chlorobenzene	0.819	0.730	10.9	72	0.00
58 T	Ethyl Benzene	1.494	1.382	7.5	76	0.00
59 T	m/p-Xylene	1.255	1.140	9.2	76	0.00
60 T	o-Xylene	1.175	1.070	8.9	76	0.00
61 T	Styrene	0.716	0.684	4.5	74	0.00
62	Isopropylbenzene	1.763	1.569	11.0	74	0.00
63 T	1,1,2,2-Tetrachloroethane	0.744	0.674	9.4	75	0.00
64	n-propylbenzene	0.458	0.409	10.7	71	0.00
65	tert-Butylbenzene	1.564	1.346	13.9	73	0.00
66 T	Benzyl Chloride	0.318	0.257	19.2	51	0.00
67	sec-Butylbenzene	2.195	1.957	10.8	75	0.00
68 S	1-Bromo-4-Fluorobenzene	0.759	0.730	3.8	73	0.00
69	p-Isopropyltoluene	1.802	1.616	10.3	73	0.00
70	n-Butylbenzene	1.816	1.738	4.3	77	0.00
71	2-Chlorotoluene	1.326	1.218	8.1	75	0.00
72 T	4-Ethyltoluene	1.403	1.328	5.3	74	0.00
73 T	1,3,5-Trimethylbenzene	1.260	1.167	7.4	74	0.00
74 T	1,2,4-Trimethylbenzene	1.332	1.210	9.2	76	0.00
75 T	1,3-Dichlorobenzene	0.773	0.702	9.2	73	0.00
76 T	1,4-Dichlorobenzene	0.762	0.690	9.4	73	0.00
77 T	1,2-Dichlorobenzene	0.762	0.694	8.9	73	0.00
78 T	Hexachloro-1,3-Butadiene	0.602	0.486	19.3	73	0.00
79 T	Naphthalene	0.996	1.023	-2.7	72	0.00
80 T	Naphthalene,2-methyl-	0.246	0.264	-7.3	68	0.00
81 T	1,2,4-Trichlorobenzene	0.583	0.550	5.7	74	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0