

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL041322\
 Data File : VL038905.D
 Acq On : 13 Apr 2022 15:38
 Operator : SY/AP
 Sample : VL0413ABL02
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0413ABL02

Quant Time: Apr 25 08:28:13 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL040522AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Apr 06 06:12:53 2022
 QLast Update : Wed Apr 06 06:12:53 2022
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1208208	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	2887890	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.09	117	2513300	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	2026778	10.63	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	106.30%


Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
39) 1,2-Dichloropropane	7.18	63	779546	7.758	ppbv	# 27


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL041322\
Data File : VL038905.D
Acq On : 13 Apr 2022 15:38
Operator : SY/AP
Sample : VL0413ABL02
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0413ABL02





Quant Time: Apr 25 08:28:13 2022
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL040522AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Apr 06 06:12:53 2022
QLast Update : Wed Apr 06 06:12:53 2022
Response via : Initial Calibration








#1
 Bromochloromethane
 Concen: 10.000 ppbv
 RT: 5.67
 Instrument: MSVOA_1
 Delta R.T.:
 Lab File: ClientSampleId : VL0413ABL02
 Acq: 13 Apr 2022 15:38


Tot Ion: 49 Resp: 1208208
 Ion Ratio Lower Upper
 49 100
 128 46.5 24.1 72.3
 130 59.6 30.9 92.8



#33
 1,4-Difluorobenzene
 Concen: 10.000 ppbv
 RT: 7.18 min Scan# 1365
 Delta R.T.: 0.00 min
 Lab File: VL038905.D
 Acq: 13 Apr 2022 15:38

Tgt Ion: 114 Resp: 2887890
 Ion Ratio Lower Upper
 114 100
 63 28.0 19.8 29.8
 88 19.4 14.2 21.4





#39

1,2-Dichloropropane

Concen: 7.758 ppbv

RT: 7.18 Instrument:

Delta R.T. MSVOA_1

Lab File: ClientSampleId:

Acq: 13 Apr 2022 15:38 VL0413ABL02


Tot Ion: 63 Resp: 779546

Ion Ratio Lower Upper

63 100

41 0.1 53.0 79.6#




62 17.9 55.0 82.4#



Abundance Ion 63.10 (62.80 to 63.80): VL0

Ion 41.10 (40.80 to 41.80): VL0

Ion 62.10 (61.80 to 62.80): VL0



#55

Chlorobenzene-d5

Concen: 10.000 ppbv

RT: 12.09 min Scan# 2893

Delta R.T. 0.01 min

Lab File: VL038905.D

Acq: 13 Apr 2022 15:38


Tgt Ion: 117 Resp: 2513300

Ion Ratio Lower Upper

117 100

82 73.4 53.0 79.6



119 33.2 24.9 37.3




Abundance Ion 117.10 (116.80 to 117.80): V

Ion 82.10 (81.80 to 82.80): VL0

Ion 119.10 (118.80 to 119.80): V





#68

1-Bromo-4-Fluorobenzene

Concen: 10.631 ppbv

RT: 14.42 Instrument:

Delta R.T. MSVOA_L

Lab File: ClientSampleId :

Acq: 13 Apr 2022 15:38 VL0413ABL02


Tot Ion: 95 Resp: 2026778

Ion Ratio Lower Upper

95 100

174 63.8 52.0 78.0

175 4.3 4.0 6.0



Abundance Ion 95.10 (94.80 to 95.80): VL0

Ion 174.00 (173.70 to 174.70): V

Ion 175.00 (174.70 to 175.70): V

