

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041322\
 Data File : VL038909.D
 Acq On : 13 Apr 2022 18:54
 Operator : SY/AP
 Sample : N2383-01DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 1DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/14/2022
 Supervised By :Mahesh Dadoda 04/18/2022

Quant Time: Apr 14 08:26:23 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL040522AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Apr 0
 QLast Update : Wed Apr 06 06:12:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.677	49	1162899	10.000	ppbv	0.01
33) 1,4-Difluorobenzene	7.182	114	2696266	10.000	ppbv	0.00
55) Chlorobenzene-d5	12.092	117	2338475	10.000	ppbv	0.01
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.420	95	1918819	10.818	ppbv	0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	108.200%
Target Compounds						
					Qvalue	
3) Chlorodifluoromethane	2.999	51	17772	0.125	ppbv	95
9) Propene	3.031	41	8303	0.111	ppbv	92
13) Ethanol	3.607	45	261168m	57.233	ppbv	
15) Acetone	3.864	43	95825	0.828	ppbv	98
20) Methylene Chloride	4.349	84	21868m	0.521	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041322\
Data File : VL038909.D
Acq On : 13 Apr 2022 18:54
Operator : SY/AP
Sample : N2383-01DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
1DL

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 04/14/2022
Supervised By : Mahesh Dadoda 04/18/2022

Quant Time: Apr 14 08:26:23 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL040522AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Apr
QLast Update : Wed Apr 06 06:12:53 2022
Response via : Initial Calibration

