

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041922\
 Data File : VL038947.D
 Acq On : 19 Apr 2022 13:12
 Operator : SY/AP
 Sample : QC041822 CAN 10439
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC041822 CAN 10439

Quant Time: Apr 20 04:25:49 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL040522AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Apr 06 06:12:53 2022
 QLast Update : Wed Apr 06 06:12:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.681	49	1151239	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	7.188	114	2996512	10.000	ppbv	0.02
55) Chlorobenzene-d5	12.095	117	2631109	10.000	ppbv	0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.423	95	2066677	10.355	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041922\
Data File : VL038947.D
Acq On : 19 Apr 2022 13:12
Operator : SY/AP
Sample : QC041822 CAN 10439
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC041822 CAN 10439

Quant Time: Apr 20 04:25:49 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL040522AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Apr 06 06:12:53 2022
QLast Update : Wed Apr 06 06:12:53 2022
Response via : Initial Calibration

