

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL042524\
 Data File : VL041165.D
 Acq On : 25 Apr 2024 10:22
 Operator : SY/MD
 Sample : VL0425ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0425ABL01

Quant Time: Apr 26 03:10:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL040424AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Apr 26 03:09:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.176	49	1254975	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.629	114	3103858	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.436	117	2725049	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.744	95	2112901	9.642	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.400%
Target Compounds						
					Qvalue	
15) Acetone	3.484	43	46588	0.343	ppbv #	88
20) Methylene Chloride	3.928	84	19110	0.341	ppbv #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL042524\
Data File : VL041165.D
Acq On : 25 Apr 2024 10:22
Operator : SY/MD
Sample : VL0425ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0425ABL01

Quant Time: Apr 26 03:10:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL040424AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri Apr 26 03:09:28 2024
Response via : Initial Calibration

