

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL042721\
 Data File : VL036854.D
 Acq On : 27 Apr 2021 12:35
 Operator : SY/AP
 Sample : M2174-01DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA1DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/28/2021
 Supervised By :Mahesh Dadoda 04/29/2021

Quant Time: Apr 28 06:12:09 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL042721A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Tue Apr 27 16:56:35 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1477142	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.17	114	3578519	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.09	117	3116475	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	2354269	9.41	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	94.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.02	41	22937m	0.280	ppbv	
13) Ethanol	3.62	45	116779m	11.169	ppbv	
15) Acetone	3.86	43	126561	0.909	ppbv #	82
19) Isopropyl Alcohol	4.02	45	24224m	0.351	ppbv	
20) Methylene Chloride	4.35	84	258078	3.876	ppbv	98
26) Hexane	5.69	57	197980	1.244	ppbv #	44

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL042721\
Data File : VL036854.D
Acq On : 27 Apr 2021 12:35
Operator : SY/AP
Sample : M2174-01DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA1DL

Manual Integrations
APPROVED

Quant Time: Apr 28 06:12:09 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL042721
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Tue Apr 27 16:56:35 2021
Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 04/28/2021
Supervised By :Mahesh Dadoda 04/29/2021

