

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL042721\
 Data File : VL036866.D
 Acq On : 27 Apr 2021 20:30
 Operator : SY/AP
 Sample : M2172-08DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 OFFSITE-DDC-INFLUENTDL

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 04/28/2021
 Supervised By : Mahesh Dadoda 04/29/2021

Quant Time: Apr 29 16:16:27 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL042721A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Tue Apr 27 16:56:35 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1540993	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	3633944	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.10	117	3200319	10.00	ppbv	0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.43	95	2495558	9.71	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.63	45	29686m	2.722	ppbv	
15) Acetone	3.88	43	124847m	0.860	ppbv	
20) Methylene Chloride	4.35	84	133257	1.919	ppbv	96
26) Hexane	5.69	57	95363	0.575	ppbv #	46
31) cis-1,2-Dichloroethene	5.56	61	93573	0.628	ppbv	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL042721\
Data File : VL036866.D
Acq On : 27 Apr 2021 20:30
Operator : SY/AP
Sample : M2172-08DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
OFFSITE-DDC-INFLUENTDL

Quant Time: Apr 29 16:16:27 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL042721
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Tue Apr 27 16:56:35 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 04/28/2021
Supervised By :Mahesh Dadoda 04/29/2021

