

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_L\Data\VL042925\  
 Data File : VL042440.D  
 Acq On : 29 Apr 2025 11:30  
 Operator : SY/MD  
 Sample : Q1880-05DL 10X  
 Misc : 400mL/MSVOA\_L  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_L  
 ClientSampleId :  
 EBR-SC-SS-DUP01-20250421DL

Quant Time: Apr 30 01:18:47 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_L\methods\VL041725AIR.M  
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA\_L Fri Aug 2  
 QLast Update : Fri Apr 18 02:14:00 2025  
 Response via : Initial Calibration

### Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 04/30/2025  
 Supervised By : Mahesh Dadoda 04/30/2025

| Compound                    | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|--------|----------|
| -----                       |        |       |          |          |        |          |
| Internal Standards          |        |       |          |          |        |          |
| 1) Bromochloromethane       | 2.793  | 49    | 132865   | 10.000   | ppbv   | 0.00     |
| 33) 1,4-Difluorobenzene     | 3.965  | 114   | 354984   | 10.000   | ppbv   | 0.00     |
| 55) Chlorobenzene-d5        | 8.885  | 117   | 287641   | 10.000   | ppbv   | 0.00     |
| System Monitoring Compounds |        |       |          |          |        |          |
| 68) 1-Bromo-4-Fluorobenzene | 10.377 | 95    | 252848   | 10.697   | ppbv   | 0.00     |
| Spiked Amount               | 10.000 | Range | 65 - 135 | Recovery | =      | 107.000% |
| Target Compounds            |        |       |          |          |        |          |
|                             |        |       |          |          | Qvalue |          |
| 13) Ethanol                 | 1.725  | 45    | 6667     | 4.120    | ppbv # | 24       |
| 15) Acetone                 | 1.839  | 43    | 25897    | 1.191    | ppbv   | 98       |
| 20) Methylene Chloride      | 2.065  | 84    | 1577     | 0.185    | ppbv # | 61       |
| 32) 1,1,1-Trichloroethane   | 3.370  | 97    | 1557m    | 0.045    | ppbv   |          |
| 52) Tetrachloroethene       | 8.241  | 164   | 966m     | 0.065    | ppbv   |          |
| -----                       |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_L\Data\VL042925\  
Data File : VL042440.D  
Acq On : 29 Apr 2025 11:30  
Operator : SY/MD  
Sample : Q1880-05DL 10X  
Misc : 400mL/MSVOA\_L  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_L  
ClientSampleId :  
EBR-SC-SS-DUP01-20250421DL

Quant Time: Apr 30 01:18:47 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_L\methods\VL041725AIR.M  
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA\_L Fri Aug  
QLast Update : Fri Apr 18 02:14:00 2025  
Response via : Initial Calibration

Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 04/30/2025  
Supervised By :Mahesh Dadoda 04/30/2025

