

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL042925\
 Data File : VL042441.D
 Acq On : 29 Apr 2025 12:02
 Operator : SY/MD
 Sample : Q1880-07DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 EBR-P16-SC-SS01-20250422DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/30/2025
 Supervised By :Mahesh Dadoda 04/30/2025

Quant Time: Apr 30 01:19:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL041725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri Apr 18 02:14:00 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	120274	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.959	114	315854	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.882	117	264275	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.377	95	230129	10.597	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	106.000%
Target Compounds						
					Qvalue	
13) Ethanol	1.722	45	29902	20.415	ppbv	77
15) Acetone	1.836	43	43049	2.186	ppbv	97
17) tert-Butyl alcohol	2.043	59	8565	0.351	ppbv #	82
19) Isopropyl Alcohol	1.887	45	12664	1.158	ppbv #	27
34) 2-Butanone	2.561	43	6473m	0.259	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL042925\
Data File : VL042441.D
Acq On : 29 Apr 2025 12:02
Operator : SY/MD
Sample : Q1880-07DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
EBR-P16-SC-SS01-20250422DL

Quant Time: Apr 30 01:19:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL041725AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Apr 18 02:14:00 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 04/30/2025
Supervised By :Mahesh Dadoda 04/30/2025

