

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL042925\
 Data File : VL042442.D
 Acq On : 29 Apr 2025 12:33
 Operator : SY/MD
 Sample : Q1880-11DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 EBR-P02-SC-SS01-20250423DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/30/2025
 Supervised By :Mahesh Dadoda 04/30/2025

Quant Time: Apr 30 01:19:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL041725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri Apr 18 02:14:00 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.790	49	124102	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.962	114	328179	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.885	117	267319	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.380	95	233850	10.646	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	106.500%
Target Compounds						
					Qvalue	
13) Ethanol	1.725	45	22020	14.570	ppbv	81
15) Acetone	1.839	43	16108	0.793	ppbv	93
19) Isopropyl Alcohol	1.890	45	4502m	0.399	ppbv	
20) Methylene Chloride	2.062	84	1359	0.171	ppbv #	87
30) Cyclohexane	3.865	84	1863m	0.114	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL042925\
Data File : VL042442.D
Acq On : 29 Apr 2025 12:33
Operator : SY/MD
Sample : Q1880-11DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
EBR-P02-SC-SS01-20250423DL

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 04/30/2025
Supervised By :Mahesh Dadoda 04/30/2025

Quant Time: Apr 30 01:19:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL041725AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Apr 18 02:14:00 2025
Response via : Initial Calibration

