

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL042925\
 Data File : VL042446.D
 Acq On : 29 Apr 2025 14:51
 Operator : SY/MD
 Sample : Q1908-03DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 B-3DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/30/2025
 Supervised By :Mahesh Dadoda 04/30/2025

Quant Time: Apr 30 01:21:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL041725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri Apr 18 02:14:00 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.790	49	113455	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.962	114	303124	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.885	117	255681	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.380	95	226776	10.794	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	107.900%
Target Compounds						
					Qvalue	
13) Ethanol	1.725	45	672m	0.486	ppbv	
15) Acetone	1.839	43	8923	0.480	ppbv	100
20) Methylene Chloride	2.062	84	1481	0.204	ppbv #	88
26) Hexane	2.826	57	3160m	0.171	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL042925\
Data File : VL042446.D
Acq On : 29 Apr 2025 14:51
Operator : SY/MD
Sample : Q1908-03DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
B-3DL

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 04/30/2025
Supervised By :Mahesh Dadoda 04/30/2025

Quant Time: Apr 30 01:21:20 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL041725AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Apr 18 02:14:00 2025
Response via : Initial Calibration

