

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL042925\
 Data File : VL042450.D
 Acq On : 29 Apr 2025 17:04
 Operator : SY/MD
 Sample : Q1880-01DL 20X
 Misc : 400mL/MSVOA_L
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 EBR-P11-SC-SS01-20250421DL

Quant Time: Apr 30 01:23:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL041725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Apr 18 02:14:00 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.794	49	113127	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.969	114	299787	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.891	117	249502	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.387	95	219145	10.689	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	106.900%
Target Compounds						
					Qvalue	
9) Propene	1.489	41	1638	0.189	ppbv	90
13) Ethanol	1.726	45	5856	4.251	ppbv	76
15) Acetone	1.839	43	163492	8.828	ppbv	96
34) 2-Butanone	2.570	43	11098	0.468	ppbv	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL042925\
Data File : VL042450.D
Acq On : 29 Apr 2025 17:04
Operator : SY/MD
Sample : Q1880-01DL 20X
Misc : 400mL/MSVOA_L
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
EBR-P11-SC-SS01-20250421DL

Quant Time: Apr 30 01:23:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL041725AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri Apr 18 02:14:00 2025
Response via : Initial Calibration

