

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050224\
 Data File : VL041193.D
 Acq On : 2 May 2024 12:40
 Operator : SY/MD
 Sample : P2334-03DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BA-20-SS03-20240425DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/06/2024
 Supervised By :Mahesh Dadoda 05/07/2024

Quant Time: May 03 03:39:44 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL042924AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 03 03:34:47 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.195	49	922254	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.655	114	2444956	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.468	117	2215530	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.780	95	1694141	10.129	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.300%
Target Compounds						
						Qvalue
13) Ethanol	3.256	45	19464m	1.584	ppbv	
15) Acetone	3.485	43	94548	0.869	ppbv	94
20) Methylene Chloride	3.935	84	57362	1.516	ppbv	96
26) Hexane	5.214	57	63533	0.655	ppbv	# 42
29) Chloroform	5.275	83	6809m	0.051	ppbv	
32) 1,1,1-Trichloroethane	6.005	97	9984m	0.075	ppbv	
34) 2-Butanone	4.809	43	17685m	0.142	ppbv	
38) Trichloroethene	7.291	130	235310	3.078	ppbv	95
52) Tetrachloroethene	10.616	164	3782m	0.052	ppbv	
53) Toluene	9.224	91	14908	0.063	ppbv	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050224\
Data File : VL041193.D
Acq On : 2 May 2024 12:40
Operator : SY/MD
Sample : P2334-03DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BA-20-SS03-20240425DL

Quant Time: May 03 03:39:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL042924AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri May 03 03:34:47 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 05/06/2024
Supervised By : Mahesh Dadoda 05/07/2024

