

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050922\
 Data File : VL039016.D
 Acq On : 10 May 2022 12:05
 Operator : SY/AP
 Sample : N2696-01DL2 200X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV1DL2

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 05/11/2022
 Supervised By : Mahesh Dadoda 05/11/2022

Quant Time: May 11 01:46:44 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL050522AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Thu May 11 01:46:44 2022
 QLast Update : Thu May 05 10:12:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.642	49	1020907	10.000	ppbv	-0.02
33) 1,4-Difluorobenzene	7.143	114	2887323	10.000	ppbv	-0.03
55) Chlorobenzene-d5	12.050	117	2526111	10.000	ppbv	-0.03
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.375	95	1933960	9.667	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.700%
Target Compounds						
					Qvalue	
9) Propene	3.002	41	23275	0.421	ppbv	87
10) Heptane	8.092	43	834523	6.149	ppbv	100
13) Ethanol	3.597	45	12662m	3.234	ppbv	
15) Acetone	3.854	43	27830m	0.289	ppbv	
20) Methylene Chloride	4.327	84	35644	0.982	ppbv #	86
26) Hexane	5.664	57	92749	0.906	ppbv #	40
59) m/p-Xylene	12.931	91	46959m	0.151	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050922\
Data File : VL039016.D
Acq On : 10 May 2022 12:05
Operator : SY/AP
Sample : N2696-01DL2 200X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV1DL2

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/11/2022
Supervised By :Mahesh Dadoda 05/11/2022

Quant Time: May 11 01:46:44 2022

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL050522AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LThu May

QLast Update : Thu May 05 10:12:15 2022

Response via : Initial Calibration

