

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL051220\
 Data File : VL035000.D
 Acq On : 12 May 2020 23:35
 Operator : SY/AP
 Sample : L2567-06DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 AA-1DL

Manual Integrations
 APPROVED

MMDadoda
 5/13/2020 5:12:32 PM

Quant Time: May 13 06:38:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL050620AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed May 0
 QLast Update : Wed May 06 14:49:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	974258	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.16	114	2278219	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.10	117	2090278	10.00	ppbv	-0.03

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.43	95	1731789	10.24	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.40%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.58	45	8490m	0.533	ppbv	
15) Acetone	3.84	43	45802	0.356	ppbv #	87
20) Methylene Chloride	4.31	84	10825m	0.254	ppbv	
26) Hexane	5.67	57	31180	0.332	ppbv #	41

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL051220\
Data File : VL035000.D
Acq On : 12 May 2020 23:35
Operator : SY/AP
Sample : L2567-06DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
AA-1DL

Manual Integrations
APPROVED

MMDadoda
5/13/2020 5:12:32 PM

Quant Time: May 13 06:38:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL050620AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed May
QLast Update : Wed May 06 14:49:39 2020
Response via : Initial Calibration

