

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL051618\
 Data File : VL032005.D
 Acq On : 16 May 2018 17:41
 Operator : SY/AP
 Sample : VL0516ABL01
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0516ABL01

Quant Time: May 17 09:10:28 2018
 Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL050318AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Thu May 03 15:52:13 2018
 QLast Update : Thu May 03 15:52:13 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.87	49	1546113	10.00	ppbv	-0.01
33) 1,4-Difluorobenzene	7.41	114	2249389	10.00	ppbv	-0.01
55) Chlorobenzene-d5	12.37	117	2355405	10.00	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.71	95	1872120	10.25	ppbv	-0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.50%

Target Compounds						Ovalue
15) Acetone	4.00	43	74839	0.34	ppbv #	88
20) Methylene Chloride	4.51	84	22708	0.31	ppbv #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL051618\
Data File : VL032005.D
Acq On : 16 May 2018 17:41
Operator : SY/AP
Sample : VL0516ABL01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0516ABL01

Quant Time: May 17 09:10:28 2018
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL050318AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu May 03 15:52:13 2018
QLast Update : Thu May 03 15:52:13 2018
Response via : Initial Calibration

