

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL051821\
 Data File : VL036967.D
 Acq On : 18 May 2021 13:51
 Operator : SY/AP
 Sample : M2262-13 0.03
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 LOD-MDL-AIR-01-QT2-2021

Manual Integrations
 APPROVED

MMDadoda
 5/20/2021 5:55:13 PM

Quant Time: May 19 03:50:47 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL051821AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue May 18 2021
 QLast Update : Tue May 18 12:00:15 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1288137	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.15	114	3658485	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.06	117	3298389	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.39	95	2482098	10.02	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.20%

Target Compounds					Ovalue
5) Vinyl Chloride	3.25	62	3553m	0.040	ppbv
32) 1,1,1-Trichloroethane	6.48	97	9707m	0.040	ppbv
35) Carbon Tetrachloride	6.99	117	8951m	0.037	ppbv
38) Trichloroethene	7.82	130	6624m	0.049	ppbv
52) Tetrachloroethene	11.20	164	6046m	0.048	ppbv
63) 1,1,2,2-Tetrachloroethane	13.66	83	12497m	0.046	ppbv

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL051821\
Data File : VL036967.D
Acq On : 18 May 2021 13:51
Operator : SY/AP
Sample : M2262-13 0.03
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
LOD-MDL-AIR-01-QT2-2021

Manual Integrations
APPROVED

MMDadoda
5/20/2021 5:55:13 PM

Quant Time: May 19 03:50:47 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL051821AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LTue May
QLast Update : Tue May 18 12:00:15 2021
Response via : Initial Calibration

