


Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL051821\
 Data File : VL036949.D
 Acq On : 18 May 2021 1:20
 Operator : SY/AP
 Sample : BFB
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL051821AIR.M
 Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Tue May 18 12:00:15 2021
 Last Update : Tue May 18 12:00:15 2021



AutoFind: Scans 3608, 3609, 3610; Background Corrected with Scan 3585

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	21.8	217685	PASS
75	95	30	66	53.4	533034	PASS
95	95	100	100	100.0	999082	PASS
96	95	5	9	6.8	68194	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	70.7	706282	PASS
175	174	4	9	7.7	54562	PASS
176	174	93	101	96.8	683626	PASS
177	176	5	9	7.0	47776	PASS