

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL051921\
 Data File : VL036996.D
 Acq On : 19 May 2021 17:48
 Operator : SY/AP
 Sample : M2424-11DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 KIA-01DL

Manual Integrations
 APPROVED

MMDadoda
 5/24/2021 3:01:40 PM

Quant Time: May 21 17:02:11 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL051821AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue May 18 2021
 QLast Update : Tue May 18 12:00:15 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1249046	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.16	114	3696640	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.06	117	3384044	10.00	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.39	95	2441953	9.61	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.01	41	26102	0.319	ppbv	88
13) Ethanol	3.59	45	661337m	60.492	ppbv	
15) Acetone	3.84	43	270931	2.118	ppbv	92
19) Isopropyl Alcohol	3.98	45	48736m	0.709	ppbv	
20) Methylene Chloride	4.33	84	665606	9.706	ppbv	98
26) Hexane	5.67	57	608848	3.877	ppbv #	47

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL051921\
Data File : VL036996.D
Acq On : 19 May 2021 17:48
Operator : SY/AP
Sample : M2424-11DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
KIA-01DL

Manual Integrations
APPROVED

MMDadoda
5/24/2021 3:01:40 PM

Quant Time: May 21 17:02:11 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL051821AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue May
QLast Update : Tue May 18 12:00:15 2021
Response via : Initial Calibration

